

Certificate



Certificate no.

TA 50224325 01

License Holder:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd., Fu Hsing Park
Fu Hsing Hsiang, Chang Hua Hsien
506
Taiwan

Manufacturing Plant:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd., Fu Hsing Park
Fu Hsing Hsiang, Chang Hua Hsien
506
Taiwan

Test report no.: ZTW1-SCL 10034395 001

Client Reference: 1108009005-006/VTC

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Adapter)

License Fee - Units

Type Designation	: TR60MX	10
X stands for	: 05, 12, 15, 18, 19, 24, 36 or 48	1
Rated Voltage	: AC 100-240V, 47-63Hz	
Rated Current	: 1) 0.75-0.4A for model TR60M05 2) 1.5-0.7A for other models	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Ambient Temperature:	: 40°C	
Protection Class	: II	

This certificate is valid for maximum 5 years.

Appendix: 1

11


Licensed Test mark:

TÜV Rheinland Taiwan Ltd.


Signature

Date of Issue
(day/mo/yr)

14/03/2012



Dipl.-Ing. W. Hsu



Certificate



Certificate no.

TA 50224325 07

License Holder:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd., Fu Hsing Park
Fu Hsing Hsiang, Chang Hua Hsien
506
Taiwan

Manufacturing Plant:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd., Fu Hsing Park
Fu Hsing Hsiang, Chang Hua Hsien
506
Taiwan

Test report no.: ZTW1-KM 10034395 006

Client Reference: VTC/230503702-03

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Switching Power Adapter)

License Fee - Units

as page 01

Change

Test Requirement : see above

Remark: 1. The risk management and usability requirements of the standard were not addressed.
2. This certificate is valid for maximum 5 years.

Appendix: 3

Licensed Test mark:




TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

23/08/2023


Dipl.-Ing. (FH) A. Klinker

